

# CE & UKCA Declaration of Conformity

Medical devices regulations (EU 2017/745 MDR) and (UK MDR 2002)  
For Class 1 Medical Devices

**Manufacturer's name** : Van Raam Reha Bikes B.V.  
**Manufacturer's address** : Guldenweg 23  
: 7051 HT Varsseveld  
: The Netherlands  
**SRN** : NL-MF-000006693  
**Representative in UK** : Van Raam Ltd  
**Product** : Therapy Tricycle for people with a disability or a limitation  
**Device Name** : Fun2Go 2  
**Basic-UDI** : 8720299493957  
**Part number** : 387-XXXX

The product of the declaration described above complies with the Essential Requirements of: EU Medical Devices Regulations 2017/745 (MDR - Class 1) and UK Medical Devices Regulation 2002 (UK MDR 2002 Class 1). The products are labelled with the CE Mark according EU 2017/745 and UKCA-Mark according to EU Exit Regulations 2020, SI2020/1478.

If the product of declaration, as delivered, is equipped with the Silent Elektro System, it also complies with the Essential Requirements of:

**Machinery Directive** : EU CE 2006/42/EU  
: UKCA Machinery (Safety) Reg. SI 2008/1597  
**Radio Equipment** : EU CE 2014/53/EU  
: UKCA Radio Equipment Reg. SI 2017/1206  
**EMC Directive** : EU CE 2014/30/EU  
: UKCA Electromagnetic Compatibility Reg. SI 2016/1091  
**Low Voltage** : EU CE 2014/35/EU  
: UKCA Electrical Equipment (Safety) Reg. SI 2016/1101  
**RoHS** : EU CE 2011/65/EU and 2015/863/EU  
: UKCA Electrical and Electronic Equipment Reg. SI 2012/3032

*For used harmonized standards see page 3*

And is in conformity with the requirements of the following harmonized standard:

**EPAC Bicycles** : BS - NEN - EN 15194:2017 Electrically power assisted cycles  
part 4.2 Electrical requirements.

Signed for and on behalf of Van Raam Reha Bikes B.V.

Varsseveld, 24-04-2023

Place, Date of Issue

  
Signature, Name, Position

# CE & UKCA Declaration of Conformity

**Manufacturer:** Kunshan ST Electronics Co., Ltd

**Manufacturer's address:** No.198 Huatao Road Qiandeng  
Kunshan  
China

**Importer:** Van Raam Reha Bikes B.V.

**Importer's address:** Guldenweg 23  
7051 HT Varsseveld  
The Netherlands

**Product:** LI-ION BATTERY CHARGER

**Device Name:** Silent charger Li-ion 36V 6A with Rosenberger connector

**Part number:** 103-5999

This declaration of conformity is issued under the sole responsibility of the importer. The product of the declaration described above complies with the Essential Requirements of:

EMC Directive : EU CE 2014/30/EU  
: UKCA Electromagnetic Compatibility Reg. SI 2016/1091

Low Voltage : EU CE 2014/35/EU  
: UKCA Electrical Equipment (Safety) Reg. SI 2016/1101

RoHS : EU CE 2011/65/EU and 2015/863/EU  
: UKCA Electrical and Electronic Equipment Reg. SI 2012/3032

*For used harmonized standards see page 3*

The products are labelled with the CE and UKCA Mark according to these directives.

Signed for and on behalf of Van Raam Reha Bikes B.V.

Varsseveld, 24-04-2023

Place, Date of Issue

Signature, Name, Position



## Used harmonized standards


<b>Standards RED – 2014/53/EU &amp; UK SI 2017/1206</b>	<b>Name</b>
ETSI EN 301 511 V12.5.1	GSM / GPRS (Telematica)
IEC 62479:2010	Bluetooth (2.4 GHz) (Display)
ETSI EN 300 328 V2.2.2	Bluetooth (2.4 GHz) (Display)
ETSI EN 303 413 V1.1.1	GLONASS / GPS (Telematica)
EN 62209-2:2010/A1:2019 / EN 50566:2017	SAR test (Battery/ Telematica)

<b>Standards EMC &amp; ESD 2014/30/EU &amp; UK SI 2016/1091</b>	<b>Name</b>
Draft ETSI EN 301 489-52 V1.1.2 (2020-12) ETSI EN 301 489-19 V2.1.1 (2019-04) Draft ETSI EN 301 489-19 V2.2.0 (2020-09) ETSI EN 301 489-17 V3.2.4 i.c.m. ETSI EN 301 489-1 V2.2.3 (2019-11)	Telematica and bluetooth
EN 55014-1:2021 / EN 55014-2:2021 EN 61000-3-2:2019 / EN 61000-3-3:2013 +A1:2019	Battery and charger combination
EN 15194:2017 Annex C	EPAC
ISO 11452-8:2015	Immunity to magnetic fields (bike)
EN 61000-4-39:2017	Immunity to Radiated fields in close proximity (bike)

<b>Standards LVD – 2014/35/EU &amp; UK SI 2016/1101</b>	<b>Name</b>
IEC 62368-1:2018	Electrical safety

Varsseveld, 24-04-2023

Place, Date of Issue



Signature, Name, Position